



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/623,113	07/18/2003	Brent R. Constantz	CORA-001CIP2CON2	4453
24353	7590	09/07/2004	EXAMINER	
BOZICEVIC, FIELD & FRANCIS LLP			BARRETT, THOMAS C	
200 MIDDLEFIELD RD			ART UNIT	
SUITE 200			PAPER NUMBER	
MENLO PARK, CA 94025			3738	

DATE MAILED: 09/07/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/623,113

Applicant(s)

CONSTANTZ, BRENT R.

Examiner

Thomas C. Barrett

Art Unit

3738

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on ____.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-5 and 16-18 is/are pending in the application.
- 4a) Of the above claim(s) ____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) ____ is/are allowed.
- 6) ☒ Claim(s) 1-5, 16-18 is/are rejected.
- 7) ☐ Claim(s) ____ is/are objected to.
- 8) ☐ Claim(s) ____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on ____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. ____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|---|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. ____. |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date <u>10-20-03</u> . | 6) <input type="checkbox"/> Other: ____. |

DETAILED ACTION

Specification

The lengthy specification has not been checked to the extent necessary to determine the presence of all possible minor errors. Applicant's cooperation is requested in correcting any errors of which applicant may become aware in the specification.

Double Patenting

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. See *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and, *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent is shown to be commonly owned with this application. See 37 CFR 1.130(b).

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claims 1-5 are rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1-4 of U.S. Patent No. 6,379,345. Although the conflicting claims are not identical, they are not patentably distinct from each other because 6,379,345 claims flushing a vascular lesion with an acidic dissolution fluid.

Claim 16 is rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claim 4 of U.S. Patent No. 6,622,732 B1. Although the conflicting claims are not identical, they are not patentably distinct from each other because 6,387,071 claims a flushing system comprising a catheter two pumping means and a dissolution fluid and the present claims are broader.

Claims 17 and 18 are rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 5 and 6 U.S. Patent No. 6,387,071. Although the conflicting claims are not identical, they are not patentably distinct from each other because 6,387,071 claims a kit and the present claims are broader.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) The invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-3, and 16-17 are rejected under 35 U.S.C. 102(b) as being anticipated by Wright (5,135,484). Wright discloses a method of treating a host suffering from a vascular disease comprising demineralizing vessels using a double balloon catheter (fig. 3) to render a local environment substantially bloodless and an acidic dissolution (col. 6, lines 33-35).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 4 and 5 are rejected under 35 U.S.C. 103(a) as being unpatentable over Wright (5,135,484) in view of Briskin (5,735,811). Wright discloses a method using a catheter for demineralization however fails to disclose the use of an energy source with the catheter. Briskin teaches a catheter with a vibrational transducer (energy source) which creates turbulence and mechanical disruption, thereby improving the effectiveness of the dissolving agent (Fig. 9). It would have been obvious to one of ordinary skill in the art to combine the teaching of a catheter with a vibrational transducer, as taught by Briskin, to a catheter for demineralization as per Wright, in order to improve the effectiveness of the dissolving agents.

Claim 18 is rejected under 35 U.S.C. 103(a) as being unpatentable over Wright (5,135,484). Wright discloses a kit comprising fluid capable of increasing proton concentration however Wright fails to disclose instructions with the kit. It is well known and obvious to one of ordinary skill in the art to combine the teaching of a kit with instructions for the kit's use, to a catheter for demineralization as per Wright, in order to ensure proper use of the kit. Please Note: a kit is a set of articles or implements used for a specific purpose. The acid and the catheter of Wright constitute a set of articles or implements used for a specific purpose and are therefore a kit. All that the printed

matter of the instructions does is teach a new use for an existing product. Where the printed matter is not functionally related to the substrate, the printed matter will not distinguish the invention from the prior art in terms of patentability.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Thomas C. Barrett whose telephone number is (703) 308-8295. The examiner can normally be reached Tuesday-Friday between 9:00 A.M. and 6:00 P.M.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott can be reached on (703) 308-2111. The fax phone numbers for the organization where this application or proceeding is assigned are (703) 872-9306 for regular communications.

Any inquiry of a general nature or relating to the status of this application or proceeding should be directed to the receptionist whose telephone number is (703) 308-0858.

Art Unit: 3738

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

A handwritten signature in black ink, appearing to read 'Thomas Barrett', with a stylized flourish at the end.

Thomas Barrett